

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
 Data File : VX037136.D
 Acq On : 21 Aug 2023 17:20
 Operator : JC/MD
 Sample : 04086-06ME
 Misc : 5.01g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-PP09BME

Quant Time: Aug 22 05:58:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	234686	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	110823	55.402	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.800%
35) Dibromofluoromethane	5.391	113	83015	53.411	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.820%
50) Toluene-d8	8.647	98	326469	57.906	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	115.820%#
62) 4-Bromofluorobenzene	11.079	95	132889	59.978	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.960%
Target Compounds						
						Qvalue
39) Methylcyclohexane	7.373	83	4379	1.887	ug/l	98
78) n-propylbenzene	11.305	91	9431	1.002	ug/l	98
89) n-Butylbenzene	12.329	91	8432	1.504	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
Data File : VX037136.D
Acq On : 21 Aug 2023 17:20
Operator : JC/MD
Sample : 04086-06ME
Misc : 5.01g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-PP09BME

Quant Time: Aug 22 05:58:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

