

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
 Data File : VX037143.D
 Acq On : 21 Aug 2023 20:07
 Operator : JC/MD
 Sample : 04014-04 10X
 Misc : 7.77g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NEVINS-SB06-Z84-85

Quant Time: Aug 22 06:02:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/22/2023
 Supervised By :Mahesh Dadoda 08/22/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	270275	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	123996	58.356	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 116.720%			
35) Dibromofluoromethane	5.385	113	93821	52.415	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.820%			
50) Toluene-d8	8.647	98	344090	52.995	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.980%			
62) 4-Bromofluorobenzene	11.079	95	135325	53.035	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.080%			
Target Compounds					Qvalue	
39) Methylcyclohexane	7.379	83	2952	1.105	ug/l	91
40) Benzene	6.038	78	1457866	199.886	ug/l	99
52) Toluene	8.720	92	27156	5.635	ug/l	99
67) Ethyl Benzene	10.195	91	4818861	515.966	ug/l	92
68) m/p-Xylenes	10.299	106	1068708	296.313	ug/l	96
69) o-Xylene	10.640	106	595319	167.300	ug/l	97
73) Isopropylbenzene	10.964	105	500645	54.988	ug/l	98
78) n-propylbenzene	11.305	91	236283	23.306	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	435316	57.317	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	1669552	217.509	ug/l	94
85) sec-Butylbenzene	11.890	105	12177m	1.439	ug/l	
86) p-Isopropyltoluene	12.012	119	147851	20.301	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
Data File : VX037143.D
Acq On : 21 Aug 2023 20:07
Operator : JC/MD
Sample : 04014-04 10X
Misc : 7.77g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NEVINS-SB06-Z84-85

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/22/2023
Supervised By :Mahesh Dadoda 08/22/2023

Quant Time: Aug 22 06:02:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

