

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082124\
 Data File : VX043075.D
 Acq On : 21 Aug 2024 09:57
 Operator : JC/MD
 Sample : VX0821MBL01
 Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0821MBL01

Quant Time: Aug 22 03:27:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 04:22:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	72762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152687	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125484	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69558	72.752	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 145.500%#	
35) Dibromofluoromethane	5.385	113	57705	61.091	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 122.180%	
50) Toluene-d8	8.646	98	170698	46.814	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.620%	
62) 4-Bromofluorobenzene	11.079	95	59004	44.628	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 89.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082124\
Data File : VX043075.D
Acq On : 21 Aug 2024 09:57
Operator : JC/MD
Sample : VX0821MBL01
Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0821MBL01

Quant Time: Aug 22 03:27:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 04:22:11 2024
Response via : Initial Calibration

