

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082219\
 Data File : VX011760.D
 Acq On : 22 Aug 2019 14:34
 Operator : JC/SP
 Sample : K4388-26
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T4-B4-(2)

Quant Time: Aug 23 02:19:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	452560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	627820	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	559331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	264796	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	213573	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
35) Dibromofluoromethane	5.49	113	185485	45.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.38%	
50) Toluene-d8	8.71	98	710869	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
62) 4-Bromofluorobenzene	11.13	95	244783	43.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.46%	
Target Compounds						
16) Acetone	2.43	43	11043	5.584	ug/l	95
20) Methylene Chloride	2.85	84	5755	1.423	ug/l	92
43) Isopropyl Acetate	6.44	43	73586	8.726	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082219\
Data File : VX011760.D
Acq On : 22 Aug 2019 14:34
Operator : JC/SP
Sample : K4388-26
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T4-B4-(2)

Quant Time: Aug 23 02:19:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

