

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082219\
 Data File : VX011765.D
 Acq On : 22 Aug 2019 16:32
 Operator : JC/SP
 Sample : K4388-34
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T4-E-(0)

Quant Time: Aug 23 02:31:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	410324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	601285	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	552131	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	263389	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	209808	50.90	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.80%	
35) Dibromofluoromethane	5.49	113	177431	45.63	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.26%	
50) Toluene-d8	8.71	98	687961	50.73	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.46%	
62) 4-Bromofluorobenzene	11.13	95	239976	44.77	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	10808	6.028	ug/l	98
20) Methylene Chloride	2.85	84	8717	2.378	ug/l	94
43) Isopropyl Acetate	6.44	43	57340	7.100	ug/l	99
95) Naphthalene	13.83	128	35688	2.167	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082219\
Data File : VX011765.D
Acq On : 22 Aug 2019 16:32
Operator : JC/SP
Sample : K4388-34
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T4-E-(0)

Quant Time: Aug 23 02:31:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

