

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082219\
 Data File : VX011780.D
 Acq On : 22 Aug 2019 22:24
 Operator : JC/SP
 Sample : K4403-25
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T1-H1-(2.5)

Quant Time: Aug 23 03:49:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	451706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	624005	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	572211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	280087	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	209272	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
35) Dibromofluoromethane	5.49	113	189371	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
50) Toluene-d8	8.71	98	721026	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
62) 4-Bromofluorobenzene	11.13	95	249040	44.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.54%	
Target Compounds						
20) Methylene Chloride	2.85	84	8073	2.000	ug/l	93
43) Isopropyl Acetate	6.43	43	66704	7.959	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082219\
Data File : VX011780.D
Acq On : 22 Aug 2019 22:24
Operator : JC/SP
Sample : K4403-25
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
T1-H1-(2.5)

Quant Time: Aug 23 03:49:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

