


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX082219\
Data File : VX011747.D
Acq On : 22 Aug 2019 08:17
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Title : SW846 8260
Last Update : Fri Aug 23 01:48:12 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	4054	PASS
75	95	30	60	49.7	12626	PASS
95	95	100	100	100.0	25397	PASS
96	95	5	9	7.1	1811	PASS
173	174	0.00	2	1.3	322	PASS
174	95	50	100	96.6	24521	PASS
175	174	5	9	7.3	1793	PASS
176	174	95	101	96.3	23608	PASS
177	176	5	9	7.0	1656	PASS