

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
 Data File : VX023875.D
 Acq On : 20 Aug 2021 14:18
 Operator : JC/MD
 Sample : VX0822WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0822WBL01

Quant Time: Aug 21 01:10:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Aug 03 12:05:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	33291	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	179610	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	159263	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	80778	31.376	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 104.600%	
60) 4-Bromofluorobenzene	11.086	95	70779	27.133	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 90.433%	
63) Toluene-d8	8.653	98	217702	30.372	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 101.233%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
Data File : VX023875.D
Acq On : 20 Aug 2021 14:18
Operator : JC/MD
Sample : VX0822WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0822WBL01

Quant Time: Aug 21 01:10:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Aug 03 12:05:42 2021
Response via : Initial Calibration

