

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
 Data File : VX023885.D
 Acq On : 20 Aug 2021 18:20
 Operator : JC/MD
 Sample : M3483-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-340-342

Quant Time: Aug 21 01:14:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143417	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	248679	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	102924	51.420	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.840%
35) Dibromofluoromethane	5.397	113	79490	48.295	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.653	98	304358	48.584	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.085	95	106051	45.986	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.980%
Target Compounds						
16) Acetone	2.392	43	3837	4.287	ug/l	95
17) Carbon Disulfide	2.514	76	1569	0.478	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
Data File : VX023885.D
Acq On : 20 Aug 2021 18:20
Operator : JC/MD
Sample : M3483-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-340-342

Quant Time: Aug 21 01:14:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

