

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
 Data File : VX023888.D
 Acq On : 20 Aug 2021 19:31
 Operator : JC/MD
 Sample : M3511-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 21 01:15:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231630	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80927	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96568	50.966	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.940%
35) Dibromofluoromethane	5.397	113	76676	50.014	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.020%
50) Toluene-d8	8.653	98	282948	48.491	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.980%
62) 4-Bromofluorobenzene	11.085	95	92635	43.125	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
Data File : VX023888.D
Acq On : 20 Aug 2021 19:31
Operator : JC/MD
Sample : M3511-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 21 01:15:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

