

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
 Data File : VX023897.D
 Acq On : 20 Aug 2021 23:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 21 01:16:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.528	0.507	4.0	82	0.00
3 P	Chloromethane	0.579	0.539	6.9	82	0.00
4 C	Vinyl Chloride	0.565	0.537	5.0#	82	0.00
5 T	Bromomethane	0.355	0.327	7.9	84	0.00
6 T	Chloroethane	0.383	0.346	9.7	83	0.00
7 T	Trichlorofluoromethane	0.892	0.883	1.0	86	0.00
8 T	Diethyl Ether	0.311	0.312	-0.3	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.487	3.0	84	0.00
10 T	Methyl Iodide	0.699	0.696	0.4	83	0.00
11 T	Tert butyl alcohol	0.141	0.133	5.7	81	0.00
12 CM	1,1-Dichloroethene	0.480	0.471	1.9#	86	0.00
13 T	Acrolein	0.084	0.085	-1.2	89	0.00
14 T	Allyl chloride	0.884	0.852	3.6	84	0.00
15 T	Acrylonitrile	0.314	0.312	0.6	86	0.00
16 T	Acetone	0.312	0.282	9.6	82	0.00
17 T	Carbon Disulfide	1.144	1.184	-3.5	87	0.00
18 T	Methyl Acetate	0.779	0.783	-0.5	90	0.00
19 T	Methyl tert-butyl Ether	1.670	1.708	-2.3	88	0.00
20 T	Methylene Chloride	0.596	0.545	8.6	85	0.00
21 T	trans-1,2-Dichloroethene	0.513	0.497	3.1	85	0.00
22 T	Diisopropyl ether	1.761	1.747	0.8	85	0.00
23 T	Vinyl Acetate	1.425	1.437	-0.8	85	0.00
24 P	1,1-Dichloroethane	0.980	0.956	2.4	85	0.00
25 T	2-Butanone	0.473	0.461	2.5	84	0.00
26 T	2,2-Dichloropropane	0.805	0.668	17.0	72	0.00
27 T	cis-1,2-Dichloroethene	0.610	0.594	2.6	85	0.00
28 T	Bromochloromethane	0.446	0.425	4.7	81	0.00
29 T	Tetrahydrofuran	0.306	0.301	1.6	85	0.00
30 C	Chloroform	1.038	1.009	2.8#	86	0.00
31 T	Cyclohexane	0.897	0.863	3.8	84	0.00
32 T	1,1,1-Trichloroethane	0.889	0.899	-1.1	86	0.00
33 S	1,2-Dichloroethane-d4	0.698	0.632	9.5	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.331	0.298	10.0	80	0.00
36 T	1,1-Dichloropropene	0.463	0.455	1.7	84	0.00
37 T	Ethyl Acetate	0.569	0.547	3.9	86	0.00
38 T	Carbon Tetrachloride	0.483	0.495	-2.5	87	0.00
39 T	Methylcyclohexane	0.564	0.549	2.7	82	0.00
40 TM	Benzene	1.369	1.324	3.3	84	0.00
41 T	Methacrylonitrile	0.309	0.304	1.6	86	0.00
42 TM	1,2-Dichloroethane	0.538	0.539	-0.2	86	0.00
43 T	Isopropyl Acetate	0.890	0.899	-1.0	88	0.00
44 TM	Trichloroethene	0.378	0.361	4.5	84	0.00
45 C	1,2-Dichloropropane	0.364	0.353	3.0#	83	0.00
46 T	Dibromomethane	0.253	0.254	-0.4	86	0.00
47 T	Bromodichloromethane	0.458	0.473	-3.3	86	0.00
48 T	Methyl methacrylate	0.434	0.444	-2.3	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
 Data File : VX023897.D
 Acq On : 20 Aug 2021 23:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 21 01:16:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	75	0.00
50 S	Toluene-d8	1.260	1.110	11.9	78	0.00
51 T	4-Methyl-2-Pentanone	0.576	0.593	-3.0	86	0.00
52 CM	Toluene	0.887	0.875	1.4#	84	0.00
53 T	t-1,3-Dichloropropene	0.521	0.537	-3.1	85	0.00
54 T	cis-1,3-Dichloropropene	0.555	0.557	-0.4	84	0.00
55 T	1,1,2-Trichloroethane	0.365	0.364	0.3	86	0.00
56 T	Ethyl methacrylate	0.569	0.596	-4.7	86	0.00
57 T	1,3-Dichloropropane	0.606	0.606	0.0	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.294	0.288	2.0	84	0.00
59 T	2-Hexanone	0.442	0.455	-2.9	84	0.00
60 T	Dibromochloromethane	0.353	0.388	-9.9	88	0.00
61 T	1,2-Dibromoethane	0.386	0.395	-2.3	85	0.00
62 S	4-Bromofluorobenzene	0.464	0.426	8.2	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.399	0.397	0.5	87	0.00
65 PM	Chlorobenzene	1.043	1.012	3.0	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.389	-4.9	87	0.00
67 C	Ethyl Benzene	1.822	1.815	0.4#	84	0.00
68 T	m/p-Xylenes	0.694	0.702	-1.2	85	0.00
69 T	o-Xylene	0.682	0.688	-0.9	84	0.00
70 T	Styrene	1.111	1.161	-4.5	85	0.00
71 P	Bromoform	0.263	0.295	-12.2	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.814	3.876	-1.6	85	0.00
74 T	N-amyl acetate	1.724	1.755	-1.8	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.250	1.185	5.2	84	0.00
76 T	1,2,3-Trichloropropane	1.166	1.145	1.8	84	0.00
77 T	Bromobenzene	0.939	0.924	1.6	83	0.00
78 T	n-propylbenzene	4.362	4.398	-0.8	83	0.00
79 T	2-Chlorotoluene	2.636	2.679	-1.6	85	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.234	-3.2	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.384	3.3	80	0.00
82 T	4-Chlorotoluene	2.993	3.052	-2.0	84	0.00
83 T	tert-Butylbenzene	3.044	3.121	-2.5	84	0.00
84 T	1,2,4-Trimethylbenzene	3.125	3.193	-2.2	84	0.00
85 T	sec-Butylbenzene	3.854	4.031	-4.6	84	0.00
86 T	p-Isopropyltoluene	3.172	3.331	-5.0	83	0.00
87 T	1,3-Dichlorobenzene	1.699	1.687	0.7	84	0.00
88 T	1,4-Dichlorobenzene	1.695	1.677	1.1	84	0.00
89 T	n-Butylbenzene	2.832	2.885	-1.9	82	0.00
90 T	Hexachloroethane	0.488	0.548	-12.3	86	0.00
91 T	1,2-Dichlorobenzene	1.646	1.609	2.2	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.281	-9.3	88	0.00
93 T	1,2,4-Trichlorobenzene	0.991	1.002	-1.1	84	0.00
94 T	Hexachlorobutadiene	0.442	0.417	5.7	82	0.00
95 T	Naphthalene	3.231	3.531	-9.3	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
Data File : VX023897.D
Acq On : 20 Aug 2021 23:03
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 21 01:16:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.987	1.013	-2.6	84	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6