



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082221\
 Data File : VX023872.D
 Acq On : 20 Aug 2021 12:44
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Tue Aug 03 12:05:42 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	10077	PASS
75	95	30	60	52.1	26168	PASS
95	95	100	100	100.0	50256	PASS
96	95	5	9	6.6	3342	PASS
173	174	0.00	2	1.1	464	PASS
174	95	50	100	81.5	40960	PASS
175	174	5	9	7.5	3082	PASS
176	174	95	101	96.1	39347	PASS
177	176	5	9	6.5	2562	PASS