

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082222\
 Data File : VX030714.D
 Acq On : 22 Aug 2022 14:46
 Operator : JC/MD
 Sample : VIBLK675
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK675

Quant Time: Aug 23 02:07:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 23 02:05:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	284335	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	265935	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	121329	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	70086	38.383	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.760%
7) Chloroethane-d5	1.666	69	50681	39.373	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.740%
11) 1,1-Dichloroethene-d2	2.300	63	107261	31.054	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.100%
21) 2-Butanone-d5	4.471	46	172209	101.788	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.790%
24) Chloroform-d	5.056	84	182943	48.671	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.340%
26) 1,2-Dichloroethane-d4	5.952	65	118501	46.431	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.860%
32) Benzene-d6	5.977	84	367288	48.216	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.440%
36) 1,2-Dichloropropane-d6	7.306	67	126509	51.187	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.380%
41) Toluene-d8	8.647	98	321225	44.736	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.480%
43) trans-1,3-Dichloroprop...	8.952	79	55982	45.532	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.060%
47) 2-Hexanone-d5	9.385	63	130292	103.426	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.430%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	182638	56.253	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	112.500%
66) 1,2-Dichlorobenzene-d4	12.323	152	123696	52.563	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082222\
Data File : VX030714.D
Acq On : 22 Aug 2022 14:46
Operator : JC/MD
Sample : VIBLK675
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK675

Quant Time: Aug 23 02:07:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 23 02:05:32 2022
Response via : Initial Calibration

