

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082223\
 Data File : VX037150.D
 Acq On : 22 Aug 2023 13:19
 Operator : JC/MD
 Sample : VX0822WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0822WBL01

Quant Time: Aug 22 15:38:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	167762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	287797	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	261268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133126	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117727	48.676	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.360%
35) Dibromofluoromethane	5.385	113	94069	49.354	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	8.646	98	338052	48.895	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.800%
62) 4-Bromofluorobenzene	11.079	95	134342	49.444	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082223\
Data File : VX037150.D
Acq On : 22 Aug 2023 13:19
Operator : JC/MD
Sample : VX0822WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0822WBL01

Quant Time: Aug 22 15:38:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

