

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082223\
 Data File : VX037155.D
 Acq On : 22 Aug 2023 15:11
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Aug 22 16:29:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	156892	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.769	114	261957	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131469	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99396	43.944	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.880%
35) Dibromofluoromethane	5.397	113	79341	45.733	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.460%
50) Toluene-d8	8.653	98	307643	48.886	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.780%
62) 4-Bromofluorobenzene	11.079	95	131168	53.038	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082223\
Data File : VX037155.D
Acq On : 22 Aug 2023 15:11
Operator : JC/MD
Sample : VIBLK
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Aug 22 16:29:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

