

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082223\
 Data File : VX037156.D
 Acq On : 22 Aug 2023 17:05
 Operator : JC/MD
 Sample : 04086-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20230818

Quant Time: Aug 22 17:32:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	174184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	292349	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	286401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	151850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	112056	44.623	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.240%
35) Dibromofluoromethane	5.379	113	91377	47.195	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	8.647	98	348348	49.600	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.200%
62) 4-Bromofluorobenzene	11.079	95	150360	54.478	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.960%
Target Compounds						
					Qvalue	
52) Toluene	8.714	92	16165	3.101	ug/l	100
68) m/p-Xylenes	10.299	106	2531	0.615	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082223\
Data File : VX037156.D
Acq On : 22 Aug 2023 17:05
Operator : JC/MD
Sample : 04086-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20230818

Quant Time: Aug 22 17:32:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

