

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082225\
 Data File : VX047494.D
 Acq On : 22 Aug 2025 13:58
 Operator : JC/MD
 Sample : Q2903-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-02-081925

Quant Time: Aug 25 01:27:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	245699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	503621	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	501212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	255941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	204131	59.364	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 118.720%#	
35) Dibromofluoromethane	5.397	113	169451	51.843	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.680%	
50) Toluene-d8	8.653	98	606272	51.895	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.800%	
62) 4-Bromofluorobenzene	11.079	95	251950	58.442	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.880%	
Target Compounds						
						Qvalue
16) Acetone	2.398	43	3111	2.407	ug/l	91
17) Carbon Disulfide	2.538	76	12418	1.330	ug/l	97
73) Isopropylbenzene	10.964	105	9776	0.488	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082225\
Data File : VX047494.D
Acq On : 22 Aug 2025 13:58
Operator : JC/MD
Sample : Q2903-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-02-081925

Quant Time: Aug 25 01:27:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

