

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082225\
 Data File : VX047497.D
 Acq On : 22 Aug 2025 15:01
 Operator : JC/MD
 Sample : Q2903-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16B-081925

Quant Time: Aug 25 01:29:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	181722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	373576	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	382152	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	203331	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	151317	59.497	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 119.000%#		
35) Dibromofluoromethane	5.397	113	124041	51.161	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.320%		
50) Toluene-d8	8.653	98	431839	49.832	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.660%		
62) 4-Bromofluorobenzene	11.079	95	188045	58.802	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 117.600%		
Target Compounds						
16) Acetone	2.398	43	4702	4.919	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082225\
Data File : VX047497.D
Acq On : 22 Aug 2025 15:01
Operator : JC/MD
Sample : Q2903-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16B-081925

Quant Time: Aug 25 01:29:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

