

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082318\
 Data File : VX004217.D
 Acq On : 23 Aug 2018 11:01
 Operator : JC/MD
 Sample : J4295-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIPBLANK

Quant Time: Aug 24 05:36:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	319830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	495711	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	448655	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	259476	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	214434	51.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.88%	
35) Dibromofluoromethane	5.50	113	186224	50.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.00%	
50) Toluene-d8	8.71	98	693870	49.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.70%	
62) 4-Bromofluorobenzene	11.14	95	246641	47.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.98%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082318\
Data File : VX004217.D
Acq On : 23 Aug 2018 11:01
Operator : JC/MD
Sample : J4295-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIPBLANK

Quant Time: Aug 24 05:36:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

