

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082319\
 Data File : VX011795.D
 Acq On : 23 Aug 2019 14:06
 Operator : JC/SP
 Sample : K4388-19
 Misc : 5.77g/5mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T6-E3-(0)

Quant Time: Aug 26 02:27:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	457040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	613355	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	552735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	299886	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	206841	45.05	ug/l	0.00
Spiked Amount			Recovery	=	90.10%	
35) Dibromofluoromethane	5.49	113	167731	42.29	ug/l	0.00
Spiked Amount			Recovery	=	84.58%	
50) Toluene-d8	8.71	98	696206	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	11.13	95	253817	46.42	ug/l	0.00
Spiked Amount			Recovery	=	92.84%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082319\
Data File : VX011795.D
Acq On : 23 Aug 2019 14:06
Operator : JC/SP
Sample : K4388-19
Misc : 5.77a/5mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T6-E3-(0)

Quant Time: Aug 26 02:27:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

