

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082319\
 Data File : VX011801.D
 Acq On : 23 Aug 2019 16:27
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 26 03:25:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	483586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	639045	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	570299	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	287331	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	195194	40.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.36%	
35) Dibromofluoromethane	5.49	113	187394	45.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.70%	
50) Toluene-d8	8.71	98	701600	48.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.36%	
62) 4-Bromofluorobenzene	11.13	95	254268	44.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082319\
Data File : VX011801.D
Acq On : 23 Aug 2019 16:27
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 26 03:25:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

