

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082319\
 Data File : VX011805.D
 Acq On : 23 Aug 2019 18:01
 Operator : JC/SP
 Sample : K4513-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Aug 26 03:34:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	460330	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	616419	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	544795	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	266589	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	210334	45.48	ug/l	0.00
Spiked Amount			Recovery	=	90.96%	
35) Dibromofluoromethane	5.49	113	182896	45.88	ug/l	0.00
Spiked Amount			Recovery	=	91.76%	
50) Toluene-d8	8.71	98	695397	50.02	ug/l	0.00
Spiked Amount			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	11.13	95	241037	43.86	ug/l	0.00
Spiked Amount			Recovery	=	87.72%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082319\
Data File : VX011805.D
Acq On : 23 Aug 2019 18:01
Operator : JC/SP
Sample : K4513-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Aug 26 03:34:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

