

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082319\
 Data File : VX011810.D
 Acq On : 23 Aug 2019 19:58
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 26 03:44:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	459619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	609988	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	542004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	274064	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	197577	42.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.58%	
35) Dibromofluoromethane	5.49	113	178080	45.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.30%	
50) Toluene-d8	8.71	98	692498	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
62) 4-Bromofluorobenzene	11.13	95	246269	45.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.58%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082319\
Data File : VX011810.D
Acq On : 23 Aug 2019 19:58
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 26 03:44:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

