

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX082319\
 Data File : VX011785.D
 Acq On : 23 Aug 2019 09:43
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 02:03:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	112	0.00
2 T	Dichlorodifluoromethane	0.467	0.468	-0.2	102	0.00
3 P	Chloromethane	0.409	0.405	1.0	108	0.00
4 C	Vinyl Chloride	0.409	0.412	-0.7#	115	0.00
5 T	Bromomethane	0.225	0.192	14.7	100	0.00
6 T	Chloroethane	0.239	0.239	0.0	102	0.00
7 T	Trichlorofluoromethane	0.651	0.658	-1.1	112	0.00
8 T	Diethyl Ether	0.221	0.221	0.0	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.403	0.396	1.7	109	0.00
10 T	Methyl Iodide	0.492	0.538	-9.3	112	0.00
11 T	Tert butyl alcohol	0.099	0.086	13.1	95	0.00
12 CM	1,1-Dichloroethene	0.393	0.383	2.5#	112	0.00
13 T	Acrolein	0.038	0.042	-10.5	136	0.00
14 T	Allyl chloride	0.614	0.583	5.0	100	0.00
15 T	Acrylonitrile	0.227	0.214	5.7	101	0.00
16 T	Acetone	0.218	0.199	8.7	100	0.00
17 T	Carbon Disulfide	0.994	0.965	2.9	106	0.00
18 T	Methyl Acetate	0.593	0.559	5.7	98	0.00
19 T	Methyl tert-butyl Ether	1.261	1.205	4.4	104	0.00
20 T	Methylene Chloride	0.447	0.409	8.5	103	0.00
21 T	trans-1,2-Dichloroethene	0.424	0.407	4.0	105	0.00
22 T	Diisopropyl ether	1.191	1.152	3.3	105	0.00
23 T	Vinyl Acetate	0.986	0.939	4.8	101	0.00
24 P	1,1-Dichloroethane	0.703	0.666	5.3	104	0.00
25 T	2-Butanone	0.314	0.293	6.7	103	0.00
26 T	2,2-Dichloropropane	0.597	0.566	5.2	105	0.00
27 T	cis-1,2-Dichloroethene	0.486	0.461	5.1	107	0.00
28 T	Bromochloromethane	0.319	0.303	5.0	102	0.00
29 T	Tetrahydrofuran	0.201	0.184	8.5	97	0.00
30 C	Chloroform	0.763	0.701	8.1#	103	0.00
31 T	Cyclohexane	0.618	0.621	-0.5	107	0.00
32 T	1,1,1-Trichloroethane	0.663	0.643	3.0	103	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.455	9.4	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.323	0.321	0.6	105	0.00
36 T	1,1-Dichloropropene	0.392	0.419	-6.9	114	0.00
37 T	Ethyl Acetate	0.415	0.406	2.2	98	0.00
38 T	Carbon Tetrachloride	0.444	0.461	-3.8	110	0.00
39 T	Methylcyclohexane	0.486	0.523	-7.6	108	0.00
40 TM	Benzene	1.213	1.245	-2.6	111	0.00
41 T	Methacrylonitrile	0.241	0.232	3.7	103	0.00
42 TM	1,2-Dichloroethane	0.443	0.422	4.7	101	0.00
43 T	Isopropyl Acetate	0.672	0.648	3.6	99	0.00
44 TM	Trichloroethene	0.374	0.392	-4.8	112	0.00
45 C	1,2-Dichloropropane	0.305	0.305	0.0	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX082319\
 Data File : VX011785.D
 Acq On : 23 Aug 2019 09:43
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 02:03:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.225	0.216	4.0	103	0.00
47 T	Bromodichloromethane	0.405	0.409	-1.0	103	0.00
48 T	Methyl methacrylate	0.334	0.325	2.7	98	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	100	0.00
50 S	Toluene-d8	1.128	1.177	-4.3	105	0.00
51 T	4-Methyl-2-Pentanone	0.443	0.442	0.2	103	0.00
52 CM	Toluene	0.785	0.806	-2.7#	107	0.00
53 T	t-1,3-Dichloropropene	0.425	0.457	-7.5	106	0.00
54 T	cis-1,3-Dichloropropene	0.463	0.498	-7.6	107	0.00
55 T	1,1,2-Trichloroethane	0.327	0.334	-2.1	108	0.00
56 T	Ethyl methacrylate	0.465	0.494	-6.2	103	0.00
57 T	1,3-Dichloropropane	0.528	0.529	-0.2	111	0.00
58 T	2-Chloroethyl Vinyl ether	0.245	0.264	-7.8	106	0.00
59 T	2-Hexanone	0.340	0.341	-0.3	101	0.00
60 T	Dibromochloromethane	0.354	0.377	-6.5	107	0.00
61 T	1,2-Dibromoethane	0.351	0.362	-3.1	106	0.00
62 S	4-Bromofluorobenzene	0.446	0.434	2.7	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	111	0.00
64 T	Tetrachloroethene	0.399	0.428	-7.3	111	0.00
65 PM	Chlorobenzene	0.976	1.005	-3.0	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.366	0.386	-5.5	111	0.00
67 C	Ethyl Benzene	1.631	1.692	-3.7#	108	0.00
68 T	m/p-Xylenes	0.629	0.678	-7.8	110	0.00
69 T	o-Xylene	0.618	0.653	-5.7	110	0.00
70 T	Styrene	1.033	1.126	-9.0	112	0.00
71 P	Bromoform	0.316	0.342	-8.2	111	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	3.013	3.111	-3.3	109	0.00
74 T	N-amyl acetate	1.167	1.128	3.3	101	0.00
75 P	1,1,2,2-Tetrachloroethane	0.997	0.958	3.9	106	0.00
76 T	1,2,3-Trichloropropane	0.877	0.829	5.5	102	0.00
77 T	Bromobenzene	0.888	0.914	-2.9	112	0.00
78 T	n-propylbenzene	3.222	3.394	-5.3	107	0.00
79 T	2-Chlorotoluene	2.004	2.004	0.0	105	0.00
80 T	1,3,5-Trimethylbenzene	2.514	2.644	-5.2	108	0.00
81 T	trans-1,4-Dichloro-2-butene	0.291	0.297	-2.1	102	0.00
82 T	4-Chlorotoluene	2.311	2.339	-1.2	106	0.00
83 T	tert-Butylbenzene	2.480	2.657	-7.1	109	0.00
84 T	1,2,4-Trimethylbenzene	2.539	2.687	-5.8	109	0.00
85 T	sec-Butylbenzene	2.837	3.045	-7.3	113	0.00
86 T	p-Isopropyltoluene	2.730	3.005	-10.1	112	0.00
87 T	1,3-Dichlorobenzene	1.530	1.624	-6.1	112	0.00
88 T	1,4-Dichlorobenzene	1.548	1.603	-3.6	111	0.00
89 T	n-Butylbenzene	2.260	2.426	-7.3	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX082319\
Data File : VX011785.D
Acq On : 23 Aug 2019 09:43
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 26 02:03:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.424	0.460	-8.5	110	0.00
91 T	1,2-Dichlorobenzene	1.543	1.583	-2.6	110	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.232	0.228	1.7	104	0.00
93 T	1,2,4-Trichlorobenzene	1.135	1.199	-5.6	109	0.00
94 T	Hexachlorobutadiene	0.622	0.678	-9.0	124	0.00
95 T	Naphthalene	3.126	3.444	-10.2	113	0.00
96 T	1,2,3-Trichlorobenzene	1.140	1.250	-9.6	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5