


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX082319\
Data File : VX011784.D
Acq On : 23 Aug 2019 09:11
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Title : SW846 8260
Last Update : Fri Aug 23 01:48:12 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	5481	PASS
75	95	30	60	48.4	16466	PASS
95	95	100	100	100.0	34053	PASS
96	95	5	9	7.2	2456	PASS
173	174	0.00	2	0.9	290	PASS
174	95	50	100	98.0	33368	PASS
175	174	5	9	7.8	2594	PASS
176	174	95	101	98.3	32805	PASS
177	176	5	9	6.9	2276	PASS