

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023907.D
 Acq On : 23 Aug 2021 18:50
 Operator : JC/MD
 Sample : M3511-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 24 07:16:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	240938	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89163	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99926	51.754	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.500%
35) Dibromofluoromethane	5.397	113	78772	49.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.653	98	291594	48.042	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.080%
62) 4-Bromofluorobenzene	11.085	95	99681	44.613	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
Data File : VX023907.D
Acq On : 23 Aug 2021 18:50
Operator : JC/MD
Sample : M3511-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 24 07:16:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

