

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082321\  
 Data File : VX023909.D  
 Acq On : 23 Aug 2021 19:37  
 Operator : JC/MD  
 Sample : M3520-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-TT-VPB-RW7-TB-20210819

Quant Time: Aug 24 07:16:36 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 17 11:57:51 2021  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.562  | 168  | 137935   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.769  | 114  | 236254   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.061 | 117  | 217829   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152  | 87059    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 98709    | 51.274   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =    | 102.540% |
| 35) Dibromofluoromethane    | 5.397  | 113   | 77897    | 49.816   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 99.640%  |
| 50) Toluene-d8              | 8.653  | 98    | 288111   | 48.409   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =    | 96.820%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 98442    | 44.932   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 89.860%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082321\  
Data File : VX023909.D  
Acq On : 23 Aug 2021 19:37  
Operator : JC/MD  
Sample : M3520-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BP-TT-VPB-RW7-TB-20210819

Quant Time: Aug 24 07:16:36 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 17 11:57:51 2021  
Response via : Initial Calibration

