

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023910.D
 Acq On : 23 Aug 2021 20:00
 Operator : JC/MD
 Sample : M3520-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-400-402

Quant Time: Aug 24 07:16:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235404	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93158	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	97849	51.425	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.840%
35) Dibromofluoromethane	5.398	113	76716	49.238	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	8.653	98	285065	48.071	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.140%
62) 4-Bromofluorobenzene	11.086	95	100554	46.061	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.120%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	6501	7.641	ug/l	89
17) Carbon Disulfide	2.514	76	1155	0.370	ug/l #	86
25) 2-Butanone	4.581	43	2105	1.634	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
Data File : VX023910.D
Acq On : 23 Aug 2021 20:00
Operator : JC/MD
Sample : M3520-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-400-402

Quant Time: Aug 24 07:16:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

