

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023911.D
 Acq On : 23 Aug 2021 20:24
 Operator : JC/MD
 Sample : M3520-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-420-422

Quant Time: Aug 24 07:16:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136855	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	232132	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96114	50.320	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.640%	
35) Dibromofluoromethane	5.403	113	76965	50.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.180%	
50) Toluene-d8	8.653	98	284540	48.658	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.320%	
62) 4-Bromofluorobenzene	11.085	95	91076	42.308	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.620%	
Target Compounds						Qvalue
16) Acetone	2.392	43	4721	5.528	ug/l	92
17) Carbon Disulfide	2.514	76	1683	0.537	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
Data File : VX023911.D
Acq On : 23 Aug 2021 20:24
Operator : JC/MD
Sample : M3520-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-420-422

Quant Time: Aug 24 07:16:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

