

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023912.D
 Acq On : 23 Aug 2021 20:48
 Operator : JC/MD
 Sample : M3520-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-440-442

Quant Time: Aug 24 07:17:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	136468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234605	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99970	52.487	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.980%			
35) Dibromofluoromethane	5.397	113	78641	50.645	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.280%			
50) Toluene-d8	8.653	98	289933	49.058	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.120%			
62) 4-Bromofluorobenzene	11.085	95	100520	46.203	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.400%			
Target Compounds						
16) Acetone	2.392	43	6115	7.180	ug/l	93
17) Carbon Disulfide	2.514	76	1858	0.595	ug/l #	92
20) Methylene Chloride	2.794	84	2450	1.506	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
Data File : VX023912.D
Acq On : 23 Aug 2021 20:48
Operator : JC/MD
Sample : M3520-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-440-442

Quant Time: Aug 24 07:17:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

