

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023913.D
 Acq On : 23 Aug 2021 21:11
 Operator : JC/MD
 Sample : M3520-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-460-462

Quant Time: Aug 24 07:17:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222860	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95152	52.096	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.200%
35) Dibromofluoromethane	5.397	113	73933	50.122	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.240%
50) Toluene-d8	8.653	98	271382	48.339	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.680%
62) 4-Bromofluorobenzene	11.085	95	95323	46.123	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.240%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	8089	9.905	ug/l	# 86
25) 2-Butanone	4.593	43	2091	1.690	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
Data File : VX023913.D
Acq On : 23 Aug 2021 21:11
Operator : JC/MD
Sample : M3520-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-460-462

Quant Time: Aug 24 07:17:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

