

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023914.D
 Acq On : 23 Aug 2021 21:35
 Operator : JC/MD
 Sample : M3520-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-480-482

Quant Time: Aug 24 07:17:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	134386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	232423	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	216571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97266	51.859	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.720%
35) Dibromofluoromethane	5.397	113	76180	49.521	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.040%
50) Toluene-d8	8.653	98	285765	48.807	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.085	95	97969	45.453	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.900%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	11071	13.201	ug/l #	83
20) Methylene Chloride	2.794	84	3829	2.390	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
Data File : VX023914.D
Acq On : 23 Aug 2021 21:35
Operator : JC/MD
Sample : M3520-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-480-482

Quant Time: Aug 24 07:17:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

