

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082321\  
 Data File : VX023900.D  
 Acq On : 23 Aug 2021 14:57  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Manual Integrations  
 APPROVED

MMDadoda  
 8/24/2021 11:43:53 AM

Quant Time: Aug 24 07:14:56 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 17 11:57:51 2021  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|------------------------------|--------|-------|----------|----------|------------|----------|
| -----                        |        |       |          |          |            |          |
| Internal Standards           |        |       |          |          |            |          |
| 1) Pentafluorobenzene        | 5.562  | 168   | 227082   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.769  | 114   | 350996   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5         | 10.055 | 117   | 338072   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024 | 152   | 168225   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds  |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4    | 5.958  | 65    | 163379   | 51.550   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 78 - 117 | Recovery | = 103.100% |          |
| 35) Dibromofluoromethane     | 5.391  | 113   | 124026   | 53.387   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | = 106.780% |          |
| 50) Toluene-d8               | 8.653  | 98    | 464127   | 52.491   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 92 - 112 | Recovery | = 104.980% |          |
| 62) 4-Bromofluorobenzene     | 11.085 | 95    | 186997   | 57.449   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | = 114.900% |          |
| Target Compounds             |        |       |          |          |            | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.173  | 85    | 116399   | 48.583   | ug/l       | 96       |
| 3) Chloromethane             | 1.295  | 50    | 120853   | 45.950   | ug/l       | 100      |
| 4) Vinyl Chloride            | 1.374  | 62    | 119942   | 46.759   | ug/l       | 100      |
| 5) Bromomethane              | 1.599  | 94    | 75320    | 46.703   | ug/l       | 98       |
| 6) Chloroethane              | 1.679  | 64    | 75913    | 39.912   | ug/l       | 97       |
| 7) Trichlorofluoromethane    | 1.886  | 101   | 202040   | 49.877   | ug/l       | 99       |
| 8) Diethyl Ether             | 2.136  | 74    | 69338    | 49.016   | ug/l       | 92       |
| 9) 1,1,2-Trichlorotrifluo... | 2.331  | 101   | 115271   | 50.571   | ug/l       | 97       |
| 10) Methyl Iodide            | 2.453  | 142   | 150304   | 47.358   | ug/l       | 93       |
| 11) Tert butyl alcohol       | 2.977  | 59    | 155090   | 241.972  | ug/l       | 98       |
| 12) 1,1-Dichloroethene       | 2.319  | 96    | 105717   | 48.509   | ug/l       | 94       |
| 13) Acrolein                 | 2.239  | 56    | 96159    | 252.111  | ug/l       | 97       |
| 14) Allyl chloride           | 2.666  | 41    | 198453   | 49.413   | ug/l       | 90       |
| 15) Acrylonitrile            | 3.069  | 53    | 342143   | 239.582  | ug/l       | 97       |
| 16) Acetone                  | 2.386  | 43    | 337567   | 238.210  | ug/l       | 89       |
| 17) Carbon Disulfide         | 2.514  | 76    | 267845   | 51.531   | ug/l       | 99       |
| 18) Methyl Acetate           | 2.709  | 43    | 171453   | 48.466   | ug/l       | 92       |
| 19) Methyl tert-butyl Ether  | 3.123  | 73    | 387097   | 51.036   | ug/l       | 99       |
| 20) Methylene Chloride       | 2.794  | 84    | 123186   | 45.509   | ug/l       | 88       |
| 21) trans-1,2-Dichloroethene | 3.093  | 96    | 113827   | 48.899   | ug/l       | 94       |
| 22) Diisopropyl ether        | 3.770  | 45    | 401054   | 50.137   | ug/l       | 98       |
| 23) Vinyl Acetate            | 3.727  | 43    | 1655248  | 255.801  | ug/l #     | 93       |
| 24) 1,1-Dichloroethane       | 3.611  | 63    | 217228   | 48.796   | ug/l       | 97       |
| 25) 2-Butanone               | 4.568  | 43    | 528889   | 246.407  | ug/l       | 92       |
| 26) 2,2-Dichloropropane      | 4.483  | 77    | 196113   | 53.674   | ug/l       | 98       |
| 27) cis-1,2-Dichloroethene   | 4.495  | 96    | 134727   | 48.640   | ug/l       | 95       |
| 28) Bromochloromethane       | 4.904  | 49    | 97556    | 48.195   | ug/l       | 93       |
| 29) Tetrahydrofuran          | 5.013  | 42    | 334017   | 240.249  | ug/l       | 88       |
| 30) Chloroform               | 5.099  | 83    | 230258   | 48.848   | ug/l       | 99       |
| 31) Cyclohexane              | 5.471  | 56    | 200376   | 49.211   | ug/l       | 92       |
| 32) 1,1,1-Trichloroethane    | 5.391  | 97    | 207570   | 51.426   | ug/l       | 98       |
| 36) 1,1-Dichloropropene      | 5.696  | 75    | 171262   | 52.688   | ug/l       | 98       |
| 37) Ethyl Acetate            | 4.727  | 43    | 196673   | 49.223   | ug/l       | 95       |
| 38) Carbon Tetrachloride     | 5.684  | 117   | 182556   | 53.859   | ug/l       | 99       |
| 39) Methylcyclohexane        | 7.385  | 83    | 213361   | 53.910   | ug/l       | 96       |
| 40) Benzene                  | 6.044  | 78    | 483035   | 50.256   | ug/l       | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082321\  
 Data File : VX023900.D  
 Acq On : 23 Aug 2021 14:57  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Manual Integrations  
 APPROVED

MMDadoda  
 8/24/2021 11:43:53 AM

Quant Time: Aug 24 07:14:56 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 17 11:57:51 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.928  | 41   | 108417   | 50.053  | ug/l   | 91       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 196612   | 52.048  | ug/l   | 94       |
| 43) Isopropyl Acetate         | 6.348  | 43   | 324659   | 51.947  | ug/l   | 92       |
| 44) Trichloroethene           | 7.129  | 130  | 132772   | 50.006  | ug/l   | 100      |
| 45) 1,2-Dichloropropane       | 7.434  | 63   | 128947   | 50.523  | ug/l   | 98       |
| 46) Dibromomethane            | 7.586  | 93   | 90655    | 51.041  | ug/l   | 97       |
| 47) Bromodichloromethane      | 7.824  | 83   | 176960   | 55.074  | ug/l   | 99       |
| 48) Methyl methacrylate       | 7.696  | 41   | 159921   | 52.453  | ug/l   | 85       |
| 49) 1,4-Dioxane               | 7.665  | 88   | 62609    | 971.653 | ug/l # | 93       |
| 51) 4-Methyl-2-Pentanone      | 8.580  | 43   | 1067180  | 264.155 | ug/l   | 90       |
| 52) Toluene                   | 8.720  | 92   | 319412   | 51.285  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 207367   | 49.966  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 8.373  | 75   | 213227   | 54.766  | ug/l # | 89       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 132612   | 51.718  | ug/l   | 99       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 219117   | 50.012  | ug/l # | 82       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 220628   | 51.878  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 515618   | 249.726 | ug/l   | 97       |
| 59) 2-Hexanone                | 9.433  | 43   | 856766   | 275.892 | ug/l   | 87       |
| 60) Dibromochloromethane      | 9.525  | 129  | 141576   | 49.994  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 143002   | 52.792  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 131569   | 48.725  | ug/l   | 94       |
| 65) Chlorobenzene             | 10.080 | 112  | 358452   | 50.811  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 134121   | 53.416  | ug/l   | 100      |
| 67) Ethyl Benzene             | 10.195 | 91   | 647447   | 52.556  | ug/l   | 99       |
| 68) m/p-Xylenes               | 10.305 | 106  | 499381   | 106.425 | ug/l   | 100      |
| 69) o-Xylene                  | 10.646 | 106  | 243033   | 52.706  | ug/l   | 99       |
| 70) Styrene                   | 10.659 | 104  | 407807   | 54.286  | ug/l   | 97       |
| 71) Bromoform                 | 10.805 | 173  | 105109   | 51.459  | ug/l   | 99       |
| 73) Isopropylbenzene          | 10.964 | 105  | 653629   | 50.940  | ug/l   | 99       |
| 74) N-amyl acetate            | 10.848 | 43   | 301039   | 51.915  | ug/l   | 90       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 201853   | 48.001  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 198446m  | 50.569  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 160108   | 50.661  | ug/l   | 97       |
| 78) n-propylbenzene           | 11.305 | 91   | 764963   | 52.123  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 457548   | 51.581  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 11.457 | 105  | 555176   | 52.650  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 72791    | 54.563  | ug/l   | 91       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 529073   | 52.547  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 11.719 | 119  | 545527   | 53.268  | ug/l   | 96       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 554807   | 52.769  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.896 | 105  | 708345   | 54.632  | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 587940   | 55.096  | ug/l   | 97       |
| 87) 1,3-Dichlorobenzene       | 11.976 | 146  | 295046   | 51.601  | ug/l   | 100      |
| 88) 1,4-Dichlorobenzene       | 12.043 | 146  | 291955   | 51.196  | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.335 | 91   | 529151   | 55.536  | ug/l   | 97       |
| 90) Hexachloroethane          | 12.542 | 117  | 97821    | 52.512  | ug/l   | 93       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 277773   | 50.145  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 47061    | 47.828  | ug/l   | 87       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 173874   | 52.154  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 80596    | 54.239  | ug/l   | 100      |
| 95) Naphthalene               | 13.780 | 128  | 586141   | 47.696  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 178272   | 53.692  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082321\  
Data File : VX023900.D  
Acq On : 23 Aug 2021 14:57  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM


Quant Time: Aug 24 07:14:56 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 17 11:57:51 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.562 min Scan# 734  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 168 Resp: 227082  
Ion Ratio Lower Upper  
168 100  
99 57.9 45.8 68.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#2  
Dichlorodifluoromethane  
Concen: 48.583 ug/l  
RT: 1.173 min Scan# 14  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 85 Resp: 116399  
Ion Ratio Lower Upper  
85 100  
87 30.6 16.4 49.2









#3  
Chloromethane  
Concen: 45.950 ug/l  
RT: 1.295 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 50 Resp: 120853  
Ion Ratio Lower Upper  
50 100  
52 33.2 26.4 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#4  
Vinyl Chloride  
Concen: 46.759 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 62 Resp: 119942  
Ion Ratio Lower Upper  
62 100  
64 32.2 25.7 38.5









#5  
Bromomethane  
Concen: 46.703 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 94 Resp: 75320  
Ion Ratio Lower Upper  
94 100  
96 93.6 76.4 114.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#6  
Chloroethane  
Concen: 39.912 ug/l  
RT: 1.679 min Scan# 97  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 64 Resp: 75913  
Ion Ratio Lower Upper  
64 100  
66 33.5 25.4 38.0









#7  
Trichlorofluoromethane  
Concen: 49.877 ug/l  
RT: 1.886 min Scan# 131  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:101 Resp: 202040  
Ion Ratio Lower Upper  
101 100  
103 65.9 53.2 79.8




Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM




#8  
Diethyl Ether  
Concen: 49.016 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57


Tgt Ion: 74 Resp: 69338  
Ion Ratio Lower Upper  
74 100  
45 96.9 44.5 133.5











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 50.571 ug/l  
RT: 2.331 min Scan# 204  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57





Tgt Ion:101 Resp: 115271  
Ion Ratio Lower Upper  
101 100  
85 46.1 34.9 52.3  
151 78.3 61.4 92.2



#10  
Methyl Iodide  
Concen: 47.358 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Tgt Ion:142 Resp: 150304  
Ion Ratio Lower Upper  
142 100  
127 47.1 33.3 49.9  
141 14.5 11.4 17.0



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM







#11  
Tert butyl alcohol  
Concen: 241.972 ug/l  
RT: 2.977 min Scan# 310  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 59 Resp: 155090  
Ion Ratio Lower Upper  
59 100  
57 10.7 7.8 11.8

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#12  
1,1-Dichloroethene  
Concen: 48.509 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 96 Resp: 105717  
Ion Ratio Lower Upper  
96 100  
61 166.1 124.6 187.0  
98 63.3 50.6 76.0









#13  
Acrolein  
Concen: 252.111 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 56 Resp: 96159  
Ion Ratio Lower Upper  
56 100  
55 73.3 56.4 84.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM




#14  
Allyl chloride  
Concen: 49.413 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 41 Resp: 198453  
Ion Ratio Lower Upper  
41 100  
39 72.3 51.6 77.4  
76 32.3 30.5 45.7









#15  
Acrylonitrile  
Concen: 239.582 ug/l  
RT: 3.069 min Scan# 325  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57





Tgt Ion: 53 Resp: 342143  
Ion Ratio Lower Upper  
53 100  
52 83.8 65.3 97.9  
51 37.1 28.2 42.2



#16  
Acetone  
Concen: 238.210 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Tgt Ion: 43 Resp: 337567  
Ion Ratio Lower Upper  
43 100  
58 28.6 27.8 41.8



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM







#17  
Carbon Disulfide  
Concen: 51.531 ug/l  
RT: 2.514 min Scan# 234  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 76 Resp: 267845  
Ion Ratio Lower Upper  
76 100  
78 9.2 7.2 10.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#18  
Methyl Acetate  
Concen: 48.466 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 43 Resp: 171453  
Ion Ratio Lower Upper  
43 100  
74 23.1 21.9 32.9





#19





Methyl tert-butyl Ether  
Concen: 51.036 ug/l  
RT: 3.123 min Scan# 334  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 73 Resp: 387097  
Ion Ratio Lower Upper  
73 100  
57 21.9 17.1 25.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED




MMDadoda  
8/24/2021 11:43:53 AM




#20

Methylene Chloride  
Concen: 45.509 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 84 Resp: 123186  
Ion Ratio Lower Upper  
84 100  
49 128.9 88.6 132.8  
51 39.7 27.0 40.6  
86 65.7 50.3 75.5









#21  
trans-1,2-Dichloroethene  
Concen: 48.899 ug/l  
RT: 3.093 min Scan# 329  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 96 Resp: 113827  
Ion Ratio Lower Upper  
96 100  
61 144.2 107.7 161.5  
98 64.1 50.4 75.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#22  
Diisopropyl ether  
Concen: 50.137 ug/l  
RT: 3.770 min Scan# 440  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57


Tgt Ion: 45 Resp: 401054  
Ion Ratio Lower Upper  
45 100  
43 72.0 58.0 87.0  
87 28.4 25.4 38.0  
59 11.3 10.2 15.2





#23  
 Vinyl Acetate  
 Concen: 255.801 ug/l  
 RT: 3.727 min Scan# 433  
 Delta R.T. -0.006 min  
 Lab File: VX023900.D  
 Acq: 23 Aug 2021 14:57




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050




Tgt Ion: 43 Resp: 1655248  
 Ion Ratio Lower Upper  
 43 100  
 86 9.8 9.9 14.9#

Manual Integrations  
 APPROVED



MMDadoda  
 8/24/2021 11:43:53 AM




#24  
 1,1-Dichloroethane  
 Concen: 48.796 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.006 min  
 Lab File: VX023900.D  
 Acq: 23 Aug 2021 14:57



Tgt Ion: 63 Resp: 217228  
 Ion Ratio Lower Upper  
 63 100  
 98 6.2 3.8 11.2  
 100 4.5 2.5 7.5











#25  
2-Butanone  
Concen: 246.407 ug/l  
RT: 4.568 min Scan# 571  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 43 Resp: 528889  
Ion Ratio Lower Upper  
43 100  
72 25.1 23.7 35.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#26  
2,2-Dichloropropane  
Concen: 53.674 ug/l  
RT: 4.483 min Scan# 557  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 77 Resp: 196113  
Ion Ratio Lower Upper  
77 100  
97 22.3 11.7 35.1





#27





cis-1,2-Dichloroethene  
Concen: 48.640 ug/l  
RT: 4.495 min Scan# 559  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 96 Resp: 134727  
Ion Ratio Lower Upper  
96 100  
61 152.0 0.0 285.6  
98 64.0 0.0 129.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED




MMDadoda  
8/24/2021 11:43:53 AM




#28

Bromochloromethane  
Concen: 48.195 ug/l  
RT: 4.904 min Scan# 626  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 49 Resp: 97556  
Ion Ratio Lower Upper  
49 100  
129 2.2 0.0 4.0  
130 78.3 68.2 102.2









#29  
Tetrahydrofuran  
Concen: 240.249 ug/l  
RT: 5.013 min Scan# 644  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 42 Resp: 334017  
Ion Ratio Lower Upper  
42 100  
72 43.1 41.7 62.5  
71 40.4 38.4 57.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#30  
Chloroform  
Concen: 48.848 ug/l  
RT: 5.099 min Scan# 658  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57


Tgt Ion: 83 Resp: 230258  
Ion Ratio Lower Upper  
83 100  
85 65.6 52.0 78.0





#31  
Cyclohexane  
Concen: 49.211 ug/l  
RT: 5.471 min Scan# 719  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 56 Resp: 200376  
Ion Ratio Lower Upper  
56 100  
69 31.2 25.3 37.9  
84 79.7 71.4 107.0

Manual Integrations  
APPROVED



MMDadoda  
8/24/2021 11:43:53 AM




#32  
1,1,1-Trichloroethane  
Concen: 51.426 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57



Tgt Ion: 97 Resp: 207570  
Ion Ratio Lower Upper  
97 100  
99 64.8 51.5 77.3  
61 45.3 34.2 51.4





#33





1,2-Dichloroethane-d4  
Concen: 51.550 ug/l  
RT: 5.958 min Scan# 799  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 65 Resp: 163379  
Ion Ratio Lower Upper  
65 100  
67 50.9 0.0 105.4

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050

Manual Integrations  
APPROVED




MMDadoda  
8/24/2021 11:43:53 AM




#34

1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.769 min Scan# 932  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 114 Resp: 350996  
Ion Ratio Lower Upper  
114 100  
63 21.2 0.0 40.8  
88 16.0 0.0 32.8









#35  
Dibromofluoromethane  
Concen: 53.387 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:113 Resp: 124026  
Ion Ratio Lower Upper  
113 100  
111 102.2 81.9 122.9  
192 19.9 16.7 25.1

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#36  
1,1-Dichloropropene  
Concen: 52.688 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 75 Resp: 171262  
Ion Ratio Lower Upper  
75 100  
110 35.0 18.1 54.4  
77 30.5 25.3 37.9





#37

Ethyl Acetate

Concen: 49.223 ug/l

RT: 4.727 min Scan# 597

Delta R.T. -0.000 min

Lab File: VX023900.D

Acq: 23 Aug 2021 14:57

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 43    | Resp: | 196673 |
| Ion      | Ratio | Lower | Upper  |
| 43       | 100   |       |        |
| 61       | 13.8  | 12.8  | 19.2   |
| 70       | 10.8  | 10.5  | 15.7   |

Instrument :





MSVOA\_X

ClientSampleId :

VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM



#38

Carbon Tetrachloride

Concen: 53.859 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.006 min

Lab File: VX023900.D

Acq: 23 Aug 2021 14:57

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 117   | Resp: | 182556 |
| Ion      | Ratio | Lower | Upper  |
| 117      | 100   |       |        |
| 119      | 95.3  | 76.2  | 114.2  |
| 121      | 29.8  | 24.8  | 37.2   |





#39

Methylcyclohexane

Concen: 53.910 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX023900.D

Acq: 23 Aug 2021 14:57

Tgt Ion: 83 Resp: 213361

Ion Ratio Lower Upper

83 100

55 78.9 58.9 88.3

98 47.6 38.1 57.1

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050

Manual Integrations

APPROVED

MMDadoda

8/24/2021 11:43:53 AM



#40

Benzene

Concen: 50.256 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.006 min

Lab File: VX023900.D




Acq: 23 Aug 2021 14:57

Tgt Ion: 78 Resp: 483035


Ion Ratio Lower Upper

78 100


77 23.6 18.9 28.3











#41  
Methacrylonitrile  
Concen: 50.053 ug/l  
RT: 4.928 min Scan# 630  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57





Tgt Ion: 41 Resp: 108417  
Ion Ratio Lower Upper  
41 100  
39 58.4 42.5 63.7  
67 70.3 65.0 97.6  
52 30.2 25.3 37.9



#42  
1,2-Dichloroethane  
Concen: 52.048 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Tgt Ion: 62 Resp: 196612  
Ion Ratio Lower Upper  
62 100  
98 7.9 0.0 20.2



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM







#43  
Isopropyl Acetate  
Concen: 51.947 ug/l  
RT: 6.348 min Scan# 863  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 43 Resp: 324659  
Ion Ratio Lower Upper  
43 100  
61 18.7 18.2 27.4  
87 12.9 12.5 18.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#44  
Trichloroethene  
Concen: 50.006 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:130 Resp: 132772  
Ion Ratio Lower Upper  
130 100  
95 98.9 0.0 198.0





#45





1,2-Dichloropropane  
Concen: 50.523 ug/l  
RT: 7.434 min Scan# 1041  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050

Tgt Ion: 63 Resp: 128947  
Ion Ratio Lower Upper  
63 100  
65 33.0 25.7 38.5

Manual Integrations  
APPROVED




MMDadoda  
8/24/2021 11:43:53 AM




#46

Dibromomethane  
Concen: 51.041 ug/l  
RT: 7.586 min Scan# 1066  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 93 Resp: 90655  
Ion Ratio Lower Upper  
93 100  
95 83.9 66.8 100.2  
174 99.1 83.1 124.7









#47  
Bromodichloromethane  
Concen: 55.074 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 83 Resp: 176960  
Ion Ratio Lower Upper  
83 100  
85 63.4 51.4 77.0  
127 9.0 6.8 10.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#48  
Methyl methacrylate  
Concen: 52.453 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 41 Resp: 159921  
Ion Ratio Lower Upper  
41 100  
69 77.8 76.0 114.0  
39 63.0 44.6 67.0









#49  
1,4-Dioxane  
Concen: 971.653 ug/l  
RT: 7.665 min Scan# 1079  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 88 Resp: 62609  
Ion Ratio Lower Upper  
88 100  
43 36.7 23.2 34.8#  
58 71.5 55.0 82.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#50  
Toluene-d8  
Concen: 52.491 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 98 Resp: 464127  
Ion Ratio Lower Upper  
98 100  
100 63.4 51.9 77.9









#51  
4-Methyl-2-Pentanone  
Concen: 264.155 ug/l  
RT: 8.580 min Scan# 1229  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 43 Resp: 1067180  
Ion Ratio Lower Upper  
43 100  
58 37.8 35.4 53.0

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#52  
Toluene  
Concen: 51.285 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 92 Resp: 319412  
Ion Ratio Lower Upper  
92 100  
91 174.6 137.1 205.7









#53  
t-1,3-Dichloropropene  
Concen: 49.966 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 75 Resp: 207367  
Ion Ratio Lower Upper  
75 100  
77 30.8 24.9 37.3

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#54  
cis-1,3-Dichloropropene  
Concen: 54.766 ug/l  
RT: 8.373 min Scan# 1195  
Delta R.T. 0.001 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 75 Resp: 213227  
Ion Ratio Lower Upper  
75 100  
77 32.1 25.7 38.5  
39 56.4 35.1 52.7#





#55  
1,1,2-Trichloroethane  
Concen: 51.718 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57





Tgt Ion: 97 Resp: 132612

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 85.9  | 69.3  | 103.9 |
| 85  | 55.2  | 44.3  | 66.5  |
| 99  | 62.0  | 50.3  | 75.5  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED




MMDadoda  
8/24/2021 11:43:53 AM




#56  
Ethyl methacrylate  
Concen: 50.012 ug/l  
RT: 9.122 min Scan# 1318  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 69 Resp: 219117

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 70.5  | 46.8  | 70.2# |
| 39  | 43.3  | 24.7  | 37.1# |














#57  
1,3-Dichloropropane  
Concen: 51.878 ug/l  
RT: 9.311 min Scan# 1349  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 76 Resp: 220628  
Ion Ratio Lower Upper  
76 100  
78 32.7 26.2 39.2



#58  
2-Chloroethyl Vinyl ether  
Concen: 249.726 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. -0.007 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57


Tgt Ion: 63 Resp: 515618  
Ion Ratio Lower Upper  
63 100  
106 27.9 23.5 35.3



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM







#59  
2-Hexanone  
Concen: 275.892 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 43 Resp: 856766  
Ion Ratio Lower Upper  
43 100  
58 51.8 30.9 92.8

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#60  
Dibromochloromethane  
Concen: 49.994 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 129 Resp: 141576  
Ion Ratio Lower Upper  
129 100  
127 77.4 38.4 115.2









#61  
1,2-Dibromoethane  
Concen: 52.792 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:107 Resp: 143002  
Ion Ratio Lower Upper  
107 100  
109 93.7 75.8 113.6




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#62  
4-Bromofluorobenzene  
Concen: 57.449 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 95 Resp: 186997  
Ion Ratio Lower Upper  
95 100  
174 78.4 0.0 154.6  
176 76.0 0.0 155.8









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050

Tgt Ion:117 Resp: 338072  
Ion Ratio Lower Upper  
117 100  
82 57.6 47.4 71.2  
119 32.7 26.6 39.8




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#64  
Tetrachloroethene  
Concen: 48.725 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:164 Resp: 131569  
Ion Ratio Lower Upper  
164 100  
166 125.4 103.4 155.2  
129 99.5 70.8 106.2  
131 95.5 72.5 108.7









#65  
Chlorobenzene  
Concen: 50.811 ug/l  
RT: 10.080 min Scan# 1475  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050

Tgt Ion:112 Resp: 358452  
Ion Ratio Lower Upper  
112 100  
114 31.8 25.8 38.6




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#66  
1,1,1,2-Tetrachloroethane  
Concen: 53.416 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:131 Resp: 134121  
Ion Ratio Lower Upper  
131 100  
133 95.8 47.7 143.1  
119 64.4 32.1 96.5









#67  
Ethyl Benzene  
Concen: 52.556 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 91 Resp: 647447  
Ion Ratio Lower Upper  
91 100  
106 30.2 24.6 37.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#68  
m/p-Xylenes  
Concen: 106.425 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57


Tgt Ion: 106 Resp: 499381  
Ion Ratio Lower Upper  
106 100  
91 204.2 162.8 244.2





#69  
o-Xylene  
Concen: 52.706 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Tgt Ion:106 Resp: 243033  
Ion Ratio Lower Upper  
106 100  
91 217.7 108.1 324.2

Manual Integrations  
APPROVED



MMDadoda  
8/24/2021 11:43:53 AM




#70  
Styrene  
Concen: 54.286 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57



Tgt Ion:104 Resp: 407807  
Ion Ratio Lower Upper  
104 100  
78 53.4 40.0 60.0  
103 55.7 43.4 65.0









#71  
Bromoform  
Concen: 51.459 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:173 Resp: 105109  
Ion Ratio Lower Upper  
173 100  
175 48.7 23.9 71.8  
254 0.1 0.1 0.1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:152 Resp: 168225  
Ion Ratio Lower Upper  
152 100  
115 87.7 41.1 123.3  
150 174.9 0.0 349.0











#73  
Isopropylbenzene  
Concen: 50.940 ug/l  
RT: 10.964 min Scan# 1620  
Delta R.T. 0.001 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:105 Resp: 653629  
Ion Ratio Lower Upper  
105 100  
120 26.9 13.2 39.6




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#74  
N-ethyl acetate  
Concen: 51.915 ug/l  
RT: 10.848 min Scan# 1601  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 43 Resp: 301039  
Ion Ratio Lower Upper  
43 100  
70 40.2 39.0 58.6  
55 23.2 21.5 32.3  
61 23.1 22.1 33.1









#75  
1,1,2,2-Tetrachloroethane  
Concen: 48.001 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 83 Resp: 201853  
Ion Ratio Lower Upper  
83 100  
131 10.7 5.1 15.2  
85 64.7 31.8 95.4

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM




#76  
1,2,3-Trichloropropane  
Concen: 50.569 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 75 Resp: 198446  
Ion Ratio Lower Upper  
75 100  
77 40.4 19.1 57.5









#77  
Bromobenzene  
Concen: 50.661 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57





Tgt Ion:156 Resp: 160108  
Ion Ratio Lower Upper  
156 100  
77 161.7 82.8 248.6  
158 96.5 49.1 147.4



#78  
n-propylbenzene  
Concen: 52.123 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Tgt Ion: 91 Resp: 764963  
Ion Ratio Lower Upper  
91 100  
120 23.2 11.5 34.5



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM







#79  
2-Chlorotoluene  
Concen: 51.581 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion: 91 Resp: 457548  
Ion Ratio Lower Upper  
91 100  
126 33.8 16.6 49.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#80  
1,3,5-Trimethylbenzene  
Concen: 52.650 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57


Tgt Ion: 105 Resp: 555176  
Ion Ratio Lower Upper  
105 100  
120 47.7 23.8 71.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 54.563 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. -0.006 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Tgt Ion: 75 Resp: 72791  
Ion Ratio Lower Upper  
75 100  
53 105.0 74.5 111.7  
89 46.7 37.9 56.9

Manual Integrations  
APPROVED



MMDadoda  
8/24/2021 11:43:53 AM




#82  
4-Chlorotoluene  
Concen: 52.547 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Tgt Ion: 91 Resp: 529073  
Ion Ratio Lower Upper  
91 100  
126 29.7 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 53.268 ug/l  
RT: 11.719 min Scan# 1744  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:119 Resp: 545527  
Ion Ratio Lower Upper  
119 100  
91 60.1 27.9 83.7  
134 22.8 11.3 33.9

Manual Integrations  
APPROVED



MMDadoda  
8/24/2021 11:43:53 AM




#84  
1,2,4-Trimethylbenzene  
Concen: 52.769 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57



Tgt Ion:105 Resp: 554807  
Ion Ratio Lower Upper  
105 100  
120 43.9 22.4 67.3









#85  
sec-Butylbenzene  
Concen: 54.632 ug/l  
RT: 11.896 min Scan# 1773  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:105 Resp: 708345  
Ion Ratio Lower Upper  
105 100  
134 20.1 10.0 29.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#86  
p-Isopropyltoluene  
Concen: 55.096 ug/l  
RT: 12.012 min Scan# 1792  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:119 Resp: 587940  
Ion Ratio Lower Upper  
119 100  
134 26.3 12.9 38.7  
91 25.4 11.7 35.0









#87  
1,3-Dichlorobenzene  
Concen: 51.601 ug/l  
RT: 11.976 min Scan# 1786  
Delta R.T. 0.007 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:146 Resp: 295046  
Ion Ratio Lower Upper  
146 100  
111 40.9 20.2 60.5  
148 63.9 31.9 95.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM




#88  
1,4-Dichlorobenzene  
Concen: 51.196 ug/l  
RT: 12.043 min Scan# 1797  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:146 Resp: 291955  
Ion Ratio Lower Upper  
146 100  
111 40.9 19.6 58.8  
148 63.6 31.6 95.0








#89  
n-Butylbenzene  
Concen: 55.536 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 91 Resp: 529151  
Ion Ratio Lower Upper  
91 100  
92 52.2 27.4 82.0  
134 25.6 13.2 39.5

Manual Integrations  
APPROVED



MMDadoda  
8/24/2021 11:43:53 AM




#90  
Hexachloroethane  
Concen: 52.512 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57



Tgt Ion: 117 Resp: 97821  
Ion Ratio Lower Upper  
117 100  
201 74.2 40.4 121.2









#91  
1,2-Dichlorobenzene  
Concen: 50.145 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57

Tgt Ion:146 Resp: 277773  
Ion Ratio Lower Upper  
146 100  
111 42.7 20.8 62.2  
148 63.3 31.9 95.5

Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/24/2021 11:43:53 AM



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 47.828 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57


Tgt Ion: 75 Resp: 47061  
Ion Ratio Lower Upper  
75 100  
155 80.2 45.3 135.9  
157 100.4 57.7 173.1





#93  
1,2,4-Trichlorobenzene  
Concen: 52.154 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Instrument :  
MSVOA\_X  
ClientSampled :  
VSTDCCC050




Tgt Ion:180 Resp: 173874  
Ion Ratio Lower Upper  
180 100  
182 94.8 47.7 143.1  
145 32.3 15.0 45.0

Manual Integrations  
APPROVED



MMDadoda  
8/24/2021 11:43:53 AM




#94  
Hexachlorobutadiene  
Concen: 54.239 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Tgt Ion:225 Resp: 80596  
Ion Ratio Lower Upper  
225 100  
223 61.9 31.1 93.2  
227 62.9 31.6 95.0





#95  
Naphthalene  
Concen: 47.696 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. 0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:128 Resp: 586141  
Ion Ratio Lower Upper  
128 100  
127 12.9 10.3 15.5  
129 11.0 8.8 13.2

Manual Integrations  
APPROVED

MMDadoda  
8/24/2021 11:43:53 AM



#96  
1,2,3-Trichlorobenzene  
Concen: 53.692 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX023900.D  
Acq: 23 Aug 2021 14:57



Tgt Ion:180 Resp: 178272  
Ion Ratio Lower Upper  
180 100  
182 94.2 47.6 142.9  
145 32.5 16.1 48.3

