

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023907.D
 Acq On : 23 Aug 2021 18:50
 Operator : JC/MD
 Sample : M3511-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 24 07:16:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	240938	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89163	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99926	51.754	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.500%
35) Dibromofluoromethane	5.397	113	78772	49.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.653	98	291594	48.042	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.080%
62) 4-Bromofluorobenzene	11.085	95	99681	44.613	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
Data File : VX023907.D
Acq On : 23 Aug 2021 18:50
Operator : JC/MD
Sample : M3511-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 24 07:16:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:168 Resp: 138341
Ion Ratio Lower Upper
168 100
99 59.8 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 51.754 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50

Tgt Ion: 65 Resp: 99926
Ion Ratio Lower Upper
65 100
67 51.0 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:114 Resp: 240938
Ion Ratio Lower Upper
114 100
63 20.2 0.0 40.8
88 16.6 0.0 32.8



#35
Dibromofluoromethane
Concen: 49.396 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50

Tgt Ion:113 Resp: 78772
Ion Ratio Lower Upper
113 100
111 103.2 81.9 122.9
192 19.6 16.7 25.1









#50
Toluene-d8
Concen: 48.042 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion: 98 Resp: 291594
Ion Ratio Lower Upper
98 100
100 63.3 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 44.613 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50

Tgt Ion: 95 Resp: 99681
Ion Ratio Lower Upper
95 100
174 80.7 0.0 154.6
176 76.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tgt Ion:117 Resp: 222523
Ion Ratio Lower Upper
117 100
82 58.2 47.4 71.2
119 32.6 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023907.D
Acq: 23 Aug 2021 18:50

Tgt Ion:152 Resp: 89163
Ion Ratio Lower Upper
152 100
115 59.1 41.1 123.3
150 157.4 0.0 349.0

