



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082321\
 Data File : VX023899.D
 Acq On : 23 Aug 2021 13:00
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Title : SW846 8260
 Last Update : Tue Aug 17 11:57:51 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	13023	PASS
75	95	30	60	53.4	34272	PASS
95	95	100	100	100.0	64208	PASS
96	95	5	9	6.3	4041	PASS
173	174	0.00	2	0.9	468	PASS
174	95	50	100	81.1	52067	PASS
175	174	5	9	7.0	3644	PASS
176	174	95	101	97.0	50493	PASS
177	176	5	9	6.3	3205	PASS