

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037173.D
 Acq On : 23 Aug 2023 11:02
 Operator : JC/MD
 Sample : VX0823MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0823MBL01

Quant Time: Aug 24 05:28:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	258952	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126186	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100128	44.398	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.800%
35) Dibromofluoromethane	5.385	113	81301	47.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.820%
50) Toluene-d8	8.646	98	307041	49.357	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.720%
62) 4-Bromofluorobenzene	11.079	95	126538	51.760	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037173.D
Acq On : 23 Aug 2023 11:02
Operator : JC/MD
Sample : VX0823MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0823MBL01

Quant Time: Aug 24 05:28:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

