

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037177.D
 Acq On : 23 Aug 2023 12:41
 Operator : JC/MD
 Sample : 04024-23DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20230815DL

Quant Time: Aug 24 09:33:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	215790	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200640	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105567	70.236	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 140.480%#		
35) Dibromofluoromethane	5.385	113	79810	55.845	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 111.700%		
50) Toluene-d8	8.647	98	264587	51.039	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.080%		
62) 4-Bromofluorobenzene	11.079	95	103863	50.982	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.960%		
Target Compounds						
					Qvalue	
44) Trichloroethene	7.123	130	64976	39.351	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037177.D
Acq On : 23 Aug 2023 12:41
Operator : JC/MD
Sample : 04024-23DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20230815DL

Quant Time: Aug 24 09:33:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

