

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037178.D
 Acq On : 23 Aug 2023 13:05
 Operator : JC/MD
 Sample : 04024-24DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-20230815DL

Quant Time: Aug 24 05:30:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	181068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97686	81.308	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	162.620%	#
35) Dibromofluoromethane	5.385	113	71055	59.253	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	118.500%	
50) Toluene-d8	8.647	98	204536	47.021	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	94.040%	
62) 4-Bromofluorobenzene	11.079	95	77997	45.628	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	91.260%	
Target Compounds						
					Qvalue	
44) Trichloroethene	7.123	130	77093	55.642	ug/l	100
64) Tetrachloroethene	9.268	164	346	0.297	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037178.D
Acq On : 23 Aug 2023 13:05
Operator : JC/MD
Sample : 04024-24DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-20230815DL

Quant Time: Aug 24 05:30:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

