

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037183.D
 Acq On : 23 Aug 2023 15:05
 Operator : JC/MD
 Sample : 04025-11RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT158S1-20230815RE

Quant Time: Aug 24 05:33:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	208047	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105241	67.087	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	134.180%#
35) Dibromofluoromethane	5.385	113	78324	56.845	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.700%
50) Toluene-d8	8.647	98	238688	47.757	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.079	95	85861	43.714	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.420%
Target Compounds						
24) 1,1-Dichloroethane	3.605	63	5090	2.281	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037183.D
Acq On : 23 Aug 2023 15:05
Operator : JC/MD
Sample : 04025-11RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT158S1-20230815RE

Quant Time: Aug 24 05:33:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

