

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037186.D
 Acq On : 23 Aug 2023 16:17
 Operator : JC/MD
 Sample : 04025-05DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D2-20230814DL

Quant Time: Aug 24 05:34:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	121987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	238099	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	111602	63.459	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 126.920%#	
35) Dibromofluoromethane	5.385	113	84595	53.647	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.300%	
50) Toluene-d8	8.647	98	302363	52.861	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.720%	
62) 4-Bromofluorobenzene	11.079	95	113775	50.615	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.240%	
Target Compounds						
44) Trichloroethene	7.123	130	98113	53.852	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037186.D
Acq On : 23 Aug 2023 16:17
Operator : JC/MD
Sample : 04025-05DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D2-20230814DL

Quant Time: Aug 24 05:34:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

