

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037188.D
 Acq On : 23 Aug 2023 17:05
 Operator : JC/MD
 Sample : 04025-07DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP06-20230814DL

Quant Time: Aug 24 05:34:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	141703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233156	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	101221	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88825	43.480	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.960%
35) Dibromofluoromethane	5.385	113	70809	45.857	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.720%
50) Toluene-d8	8.647	98	273167	48.770	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.540%
62) 4-Bromofluorobenzene	11.079	95	107111	48.661	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.320%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	942	0.527	ug/l #	79
44) Trichloroethene	7.123	130	67019	37.565	ug/l	97
64) Tetrachloroethene	9.275	164	448	0.290	ug/l	85
95) Naphthalene	13.780	128	2350	0.278	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037188.D
Acq On : 23 Aug 2023 17:05
Operator : JC/MD
Sample : 04025-07DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP06-20230814DL

Quant Time: Aug 24 05:34:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

