

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037189.D
 Acq On : 23 Aug 2023 17:29
 Operator : JC/MD
 Sample : 04024-15DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20230815DL

Quant Time: Aug 24 05:35:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224244	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201414	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94600	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84710	43.231	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	86.460%		
35) Dibromofluoromethane	5.385	113	67641	45.546	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.100%		
50) Toluene-d8	8.647	98	263063	48.832	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.660%		
62) 4-Bromofluorobenzene	11.079	95	100854	47.639	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.280%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.325	101	789	0.549	ug/l	92
44) Trichloroethene	7.123	130	167771	97.775	ug/l	96
64) Tetrachloroethene	9.275	164	1147	0.781	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037189.D
Acq On : 23 Aug 2023 17:29
Operator : JC/MD
Sample : 04024-15DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20230815DL

Quant Time: Aug 24 05:35:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

