

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037191.D
 Acq On : 23 Aug 2023 18:18
 Operator : JC/MD
 Sample : 04024-19DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW17811-20230815DL

Quant Time: Aug 24 05:36:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	131255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	216395	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92785	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82571	43.636	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.280%
35) Dibromofluoromethane	5.385	113	65529	45.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.440%
50) Toluene-d8	8.647	98	253108	48.688	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.380%
62) 4-Bromofluorobenzene	11.079	95	98853	48.388	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.780%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.483	96	11789	6.511	ug/l	80
44) Trichloroethene	7.123	130	120070	72.514	ug/l	95
45) 1,2-Dichloropropane	7.427	63	7201	5.001	ug/l	94
64) Tetrachloroethene	9.275	164	5860	4.080	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037191.D
Acq On : 23 Aug 2023 18:18
Operator : JC/MD
Sample : 04024-19DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW17811-20230815DL

Quant Time: Aug 24 05:36:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

