

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037194.D
 Acq On : 23 Aug 2023 19:30
 Operator : JC/MD
 Sample : 04051-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE139D1-20230816

Quant Time: Aug 24 05:37:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	225868	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86852	43.638	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.280%
35) Dibromofluoromethane	5.385	113	70018	46.808	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.620%
50) Toluene-d8	8.647	98	266370	49.091	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.180%
62) 4-Bromofluorobenzene	11.079	95	103695	48.629	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.260%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	5156	3.533	ug/l	91
12) 1,1-Dichloroethene	2.319	96	3963	2.789	ug/l #	75
27) cis-1,2-Dichloroethene	4.489	96	2500	1.313	ug/l	84
30) Chloroform	5.099	83	1966	0.619	ug/l	92
38) Carbon Tetrachloride	5.672	117	2367	1.025	ug/l	90
44) Trichloroethene	7.123	130	290244	167.935	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	1587	0.968	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037194.D
Acq On : 23 Aug 2023 19:30
Operator : JC/MD
Sample : 04051-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE139D1-20230816

Quant Time: Aug 24 05:37:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

