

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
 Data File : VX037196.D
 Acq On : 23 Aug 2023 20:19
 Operator : JC/MD
 Sample : 04051-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D-20230816

Quant Time: Aug 24 05:38:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227204	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86419	44.173	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.340%
35) Dibromofluoromethane	5.385	113	69710	46.328	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	8.647	98	264839	48.521	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	101705	47.415	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.840%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	5297	3.047	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037196.D
Acq On : 23 Aug 2023 20:19
Operator : JC/MD
Sample : 04051-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D-20230816

Quant Time: Aug 24 05:38:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

