



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082323\
Data File : VX037171.D
Acq On : 23 Aug 2023 10:11
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Title : SW846 8260
Last Update : Fri Aug 18 09:30:38 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	9050	PASS
75	95	30	60	51.3	27301	PASS
95	95	100	100	100.0	53264	PASS
96	95	5	9	6.6	3534	PASS
173	174	0.00	2	0.9	332	PASS
174	95	50	100	72.5	38605	PASS
175	174	5	9	7.5	2879	PASS
176	174	95	101	95.2	36752	PASS
177	176	5	9	6.6	2429	PASS