

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082324\
 Data File : VX043094.D
 Acq On : 23 Aug 2024 08:59
 Operator : JC/MD
 Sample : VX0823MBL01
 Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0823MBL01

Quant Time: Aug 24 00:45:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 08:42:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	281859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	539198	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	476566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	206990	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	216509	46.731	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.460%
35) Dibromofluoromethane	5.379	113	174257	46.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	8.646	98	670358	49.968	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	220866	47.300	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082324\
Data File : VX043094.D
Acq On : 23 Aug 2024 08:59
Operator : JC/MD
Sample : VX0823MBL01
Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0823MBL01

Quant Time: Aug 24 00:45:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082224W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 08:42:59 2024
Response via : Initial Calibration

