

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
 Data File : VX004243.D
 Acq On : 24 Aug 2018 13:36
 Operator : JC/MD
 Sample : J4616-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VB27184

Quant Time: Aug 25 11:51:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	297534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	471290	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	420891	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	220334	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	201323	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
35) Dibromofluoromethane	5.50	113	176458	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
50) Toluene-d8	8.71	98	668250	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	11.14	95	221675	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
Target Compounds						
11) Tert butyl alcohol	3.06	59	81299	98.349	ug/l	100
40) Benzene	6.13	78	14340	0.906	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
Data File : VX004243.D
Acq On : 24 Aug 2018 13:36
Operator : JC/MD
Sample : J4616-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VB27184

Quant Time: Aug 25 11:51:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

