

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
 Data File : VX004246.D
 Acq On : 24 Aug 2018 14:55
 Operator : JC/MD
 Sample : J4646-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-Q4-TB01

Quant Time: Aug 25 11:53:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	323633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	505093	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	433758	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	263944	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	215194	51.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.04%	
35) Dibromofluoromethane	5.50	113	189208	50.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.70%	
50) Toluene-d8	8.71	98	705928	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
62) 4-Bromofluorobenzene	11.14	95	242284	45.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
Data File : VX004246.D
Acq On : 24 Aug 2018 14:55
Operator : JC/MD
Sample : J4646-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-Q4-TB01

Quant Time: Aug 25 11:53:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

