

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
 Data File : VX004249.D
 Acq On : 24 Aug 2018 16:15
 Operator : JC/MD
 Sample : J4652-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SONO-W

Manual Integrations
 APPROVED

MMDadoda
 8/25/2018 11:43:00 AM

Quant Time: Aug 25 12:02:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	301706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	470554	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	410622	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	260062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	203182	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
35) Dibromofluoromethane	5.50	113	177515	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
50) Toluene-d8	8.71	98	657048	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
62) 4-Bromofluorobenzene	11.14	95	247687	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	69398	36.331	ug/l	99
25) 2-Butanone	4.70	43	19344	4.546	ug/l	99
29) Tetrahydrofuran	5.16	42	11636	6.887	ug/l	99
40) Benzene	6.14	78	11856	0.750	ug/l	94
52) Toluene	8.79	92	32903	3.325	ug/l	98
67) Ethyl Benzene	10.26	91	22696	1.356	ug/l	99
68) m/p-Xylenes	10.36	106	24107	3.684	ug/l	98
69) o-Xylene	10.70	106	19285	3.115	ug/l	91
73) Isopropylbenzene	11.02	105	4937	0.289	ug/l	91
78) n-propylbenzene	11.36	91	11973	0.609	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	27211	1.924	ug/l	94
84) 1,2,4-Trimethylbenzene	11.81	105	93754	6.457	ug/l	100
85) sec-Butylbenzene	11.94	105	6963	0.412	ug/l	90
95) Naphthalene	13.83	128	28679	1.690	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
Data File : VX004249.D
Acq On : 24 Aug 2018 16:15
Operator : JC/MD
Sample : J4652-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SONO-W

Manual Integrations
APPROVED

MMDadoda
8/25/2018 11:43:00 AM

Quant Time: Aug 25 12:02:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

